

Bovine IL-10 ELISA Kit

IL-10, also known as human cytokine synthesis inhibitory factor (CSIF), is an anti-inflammatory cytokine that is produced by T cells, NK cells, mast cells and macrophages (1,2,3). It is capable of inhibiting synthesis of pro-inflammatory cytokines like IFN- γ , IL-2, IL-3, TNF α and GM-CSF made by cells such as macrophages and regulatory T-cells. IL-10 also displays potent abilities to suppress the antigen presentation capacity of antigen presenting cells. IL-10 initiates signal transduction by binding to a cell surface receptor complex consisting of IL-10 RI and IL-10 RII (1). Binding of IL-10 leads to the activation of Jak1 and Tyk2, which phosphorylates Stat-3 (1,4). The anti-inflammatory activity of IL-10 is due to its ability to block signaling through other cytokine receptors, notably IFN γ receptor, by upregulating expression of SOCS-1 (1,4). In addition, IL-10 promotes T cell tolerance by inhibiting tyrosine phosphorylation of CD28 (5,6). IL-10 is an important negative regulator of the immune response, which allows for maintenance of pregnancy (1). In contrast, increased IL-10 levels contribute to persistent *Leishmania major* infections (7).

References

1. Pestka, S. et al. (2004) *Immunol Rev* 202, 8-32.
2. Akuffo, H. et al. (1999) *Clin Exp Immunol* 117, 529-34.
3. Grimbaldston MA, et al (2007). *Nat. Immunol.* 8 (10): 1095–104.
4. O'Shea, J.J. and Murray, P.J. (2008) *Immunity* 28, 477-87.
5. Akdis, C.A. and Blaser, K. (2001) *Immunology* 103, 131-6.
6. Akdis, C.A. et al. (2000) *FASEB J* 14, 1666-8.
7. Von Stebut, E. (2000) *Eur J Dermatol* 17, 115-22.

PRINCIPLE OF THE ASSAY

This is a quick ELISA assay that reduces time to 50% compared to the conventional method, and the entire assay only takes 3 hours. This assay employs the quantitative sandwich enzyme immunoassay technique and uses biotin-streptavidin chemistry to improve the performance of the assays. An antibody specific for bovineIL-10 has been pre-coated onto a microplate. Standards and samples are pipetted into the wells and any IL-10 present is bound by the immobilized antibody. After washing away any unbound substances, a detection antibody specific for bovineIL-10 is added to the wells. Following wash to remove any unbound antibody reagent, a detection reagent is added. After intensive wash a substrate solution is added to the wells and color develops in proportion to the amount of IL-10 bound in the initial step. The color development is stopped and the intensity of the color is measured.

This package insert must be read in its entirety before using this product.

Storage

Store at 4 °C. The kit can be used in 3 months.

MATERIALS PROVIDED

Description	Quantity	Description	Quantity	Description	Quantity
Antibody Precoated Plate	1	20 x PBS	1	Substrate Solution	1
Detection Antibody	1	20 x Wash Buffer	1	Stop Solution	1
Conjugate	1	10 x Reagent Diluent	1	DataSheet	1
Standard	3	20 x Standard/Sample Diluent	1	96-well plate sheet	1

Bring all reagents to room temperature before use.

Reagent Preparations

Bovine IL-10 Detection Antibody (1 vial) – The lyophilized Detection Antibody should be stored at 4°C in a manual defrost freezer for up to 3 months, if not used immediately. Centrifuge for 1 min at 6000 x g to bring down the material prior to open the vial. The vial contains sufficient Detection Antibody for a 96-well plate. Add 200 µL of sterile 1 x PBS to the antibody vial and vortex briefly and sit for 5 min. Take 200 µL of detection antibody to 10.5 mL of 1 x Reagent Diluent to make **working dilution of Detection Antibody** if the entire 96-well plate is used. If the partial antibody is used store the rest at -20°C until use.

Bovine IL-10 Standard (3 vials) – The lyophilized bovine IL-10 Standard has a total of 3 vials. Each vial contains the standard sufficient for a 96-well plate. The non-reconstituted standard can be stored at 4°C for up to 3 months if not used immediately. Centrifuge for 1 min at 6000 x g to bring down the material prior to open the tube. Add 500 µL of 1 x Standard/Sample Diluent to one Standard vial to make the high standard concentration of 18000 pg /ml. Vortex 20 sec and allow it to sit for 5 min prior to use. A seven point standard curve is generated using 2-fold serial dilutions in the Standard/Sample Diluent, vortex 20 sec for each of dilution step.

Conjugate (55 µL) – Centrifuge for 1 min at 6000 x g to bring down the material prior to open the vial. The vial contains sufficient Conjugate for a 96-well plate. If the volume is less than 55 µL, add sterile 1 x PBS to reach 55 µL and vortex briefly. Make 1:200 dilution in 1 x Reagent Diluent. If the entire 96-well plate is used, add 55 µL of Conjugate to 11 mL of 1 x Reagent Diluent to make **working dilution of Conjugate** prior to the assay. The rest of undiluted Conjugate can be stored at 4°C for up to 3 months. DO NOT FREEZE.

20 x PBS, pH 7.3, 30 mL- Dilute to 1 x PBS with deionized distilled water and mix well prior to use.

20 x Wash Buffer, 20 mL- Dilute to 1 x Wash Buffer with 1 x PBS prior to use.

10 x Reagent Diluent -Add 3 mL of sterile 1 x PBS to make 10 x Reagent Diluent, vortex for 1 min and allow it to sit for 15 min to completely dissolve. Store at -20°C. Prior to use dilute to 1 x Reagent Diluent with 1 x PBS.

20 x Standard/Sample Diluent, 10 mL- Dilute to 1 x Standard/Sample Diluent with 1 x PBS.

Substrate Solution, 11 mL.

Stop Solution, 5.5 mL.

Assay Procedure

1. Lift the plate cover from the top left corner and cover the wells that are not used. Vortex briefly the samples prior to the assay. Add 100 μ L of **sample** (such as plasma or serum) or **standards** per well and use duplicate wells for each sample. Cover the 96-well plate and incubate 1 hour at room temperature.
2. Aspirate each well and wash with **1 x Wash Buffer**, repeating the process two times for a total of three washes. Wash by filling each well with 1 x Wash Buffer (300 μ L) using a multi-channel pipette, manifold dispenser or auto-washer. Complete removal of liquid at each step is essential for good performance. After the last wash, remove any remaining Wash Buffer by aspirating or by inverting the plate and blotting it against clean paper towels.
3. Add 100 μ L of the **working dilution of Detection Antibody** to each well. Cover the plate and incubate 1 hour at room temperature.
4. Repeat the aspiration/wash as in step 2.
5. Add 100 μ L of the **working dilution of Conjugate** to each well. Cover the plate and incubate for 20 minutes at room temperature. Avoid placing the plate in direct light.
6. Repeat the aspiration/wash as in step 2.
7. Add 100 μ L of **Substrate Solution** to each well. Incubate for 10-20 minutes at room temperature. Avoid placing the plate in direct light.
8. Add 50 μ L of **Stop Solution** to each well. Gently tap the plate to ensure thorough mixing.
9. Determine the optical density of each well immediately, using a microplate reader set to 450 nm. If wavelength correction is available, set to 540 nm or 570 nm. If wavelength correction is not available, subtract readings at 540 nm or 570 nm from the readings at 450 nm. This subtraction will correct for optical imperfections in the plate. Readings made directly at 450 nm without correction may be higher and less accurate.

Precaution and Technical Notes

1. It is critical to follow the procedure step by step otherwise appropriate color development may not occur as expected.
2. A standard curve should be generated for each set of samples assayed. Thorough mixing of the standard at each dilution step is critical to ensure the normal standard curve.
3. If IL-10 exceeds the upper limit of the detection, the sample needs to be diluted with 1 x Standard/Sample Diluent. The dilution factor must be used for calculation of the concentration.
4. Conjugate contains enzyme, DO NOT mix up with Detection Antibody.
5. The Stop Solution is an acid solution, handle with caution.
6. This kit should not be used beyond the expiration date on the label.
7. A thorough and consistent wash technique is essential for proper assay performance. Wash Buffer should be dispensed forcefully and removed completely from the wells by aspiration or decanting. Remove any remaining Wash Buffer by aspiration or by inverting the plate and blotting it against clean paper towels.
8. Use a fresh reagent reservoir and pipette tips for each step.
9. It is recommended that all standards and samples be assayed in duplicate.
10. Avoid microbial contamination of reagents and buffers. This may interfere with the sensitivity of the assay.


Calculation of Results

Average the duplicate readings for each standard, control, and sample and subtract the average zero (blank) standard optical density.

Create a standard curve by reducing the data using computer software capable of generating a four parameter logistic (4-PL) curve-fit. As an alternative, construct a standard curve by plotting the mean absorbance for each standard on the y-axis against the concentration on the x-axis and draw a best fit curve through the points on the graph. The data may be linearized by plotting the log of the IL-10 concentrations versus the log of the O.D. and the best fit line can be determined by regression analysis. This procedure will produce an adequate but less precise fit of the data. If samples have been diluted, the concentration read from the standard curve must be multiplied by the dilution factor.

The Standard Curve

The graph below represents typical data generated when using this bovine IL-10 ELISA Kit. The standard curve was calculated using a computer generated 4-PL curve-fit. For this case, a Bio-Rad iMarkTM Microplate Reader and a Microplate Manager 6 Software were used to generate this curve. The correlation coefficient (r^2) is 0.999-1.000.



Specificity

The following recombinant bovine proteins prepared at 1 ng/ml were tested and exhibited no cross-reactivity or interference.

ApoA1, BMP1, BMP2, BMP3, BMP4, CCL4/MIP-1 β , CRP, HSP27, IL-1 β , IL-1ra, IL-2, IL-4, IL-5, IL-6, IL-8, IL-12, IL-15, IL-17C, IL-21, IL-23, IL2R, IL6R, IFN γ , PDGF, PLA2G7, prolactin, TGF β 1, TGF β 2, TGF β 3, TLR1, TLR2, TLR3, TNF- α , TNF RI, TNF RII, VEGF.

Calibration

This kit is calibrated against a highly purified yeast-expressed recombinant bovine IL-10.

Detection Range

280-18000 pg/ml

Assay Sensitivity

20 pg/ml

Assay Precision

Intra-Assay %CV: 5; Inter-Assay %CV: 8

For Research Use Only.

Related products

20 x PBS
10 x ELISA Wash Buffer
10 x ELISA Reagent Diluent
Universal Blocking Buffer
2 x Recombinant Protein Stabilizer
5 x Recombinant Protein Stabilizer
ELISA G-Blue Substrate Solution
Recombinant bovine IL-10
Bovine IL-10 Detection Antibody

Troubleshooting Guide

Problem	Possible causes	Solution
Poor standard curve	<ul style="list-style-type: none"> • Inaccurate pipetting • Improper standard curve 	<ul style="list-style-type: none"> • Check pipettes • Check and use the correct dilution buffer • Vortex 30 sec for each of standard dilution steps
Low signal	<ul style="list-style-type: none"> • Improper preparation of standard, samples, detection antibody, and/or conjugate • Too brief incubation times • Inadequate reagent volume or improper dilution 	<ul style="list-style-type: none"> • Briefly spin down vials before opening. Reconstitute the powder thoroughly. • Ensure sufficient incubation time. • Check pipettes and ensure correct preparation.
Large CV	<ul style="list-style-type: none"> • Inaccurate pipetting and mixing • Improper standard/sample dilutions. • Air bubbles in wells. 	<ul style="list-style-type: none"> • Check pipettes and ensure thorough mixing. • Use the correct dilution buffers • Remove bubbles in wells.
High background	<ul style="list-style-type: none"> • Plate is insufficiently washed. • Contaminated wash buffer 	<ul style="list-style-type: none"> • Review the datasheet for proper wash. If using a plate washer, ensure that all ports are unobstructed. • Make fresh wash buffer
No signal detected	<ul style="list-style-type: none"> • The procedure was misconducted. 	<ul style="list-style-type: none"> • Ensure the step-by-step protocol was correctly followed and no misstep was conducted.